

Laboratory evaluation of portable water quality testing kits



Portable testing kit	Fluidion ALERT LAB	
Manufacturer	FLUIDION SAS Bio&D, Et 1 C.C-ial Echat 94, Avenue du Général de Gaulle 94000 Créteil France Phone: +33 182 390 290 Email: contact@fluidion.com	Fluidion US Inc. / LACI 525 S. Hewitt St. Los Angeles, CA 90013, USA Phone 626-765-5580 Email contact-us@fluidion.com
Evaluation procedure	Independent laboratory evaluation, Phase 1 and Phase 2	
JMP report issue date	October, 2022	

Summary

This report summarizes the results of an independent laboratory assessment of a portable water quality testing kit called Fluidion ALERT LAB. The evaluation was carried out at KWR Research laboratory, with support from the WHO/UNICEF Joint Monitoring Programme for Water Supply, Sanitation and Hygiene (JMP) following a protocol established by WHO. The Fluidion ALERT LAB successfully passed through Phase 1 testing and proceeded to the Phase assessment, including analysis of five natural water matrices labelled N1-N5, with the following results:

- No false positives were found due to non-target bacteria, and no false negatives were found due to competition
- The portable kit results were compared in triplicate against a reference method using five different natural water matrices, and four different levels of *E. coli* contamination. This report presents the results quantitatively and qualitatively.
- Across the four test waters (excluding sterilized blanks) and five natural water matrices, a total of 60 paired samples were tested. In 87% of these, the semi-quantitative risk class matched the expected value. Matches were lowest for natural water N1 (75%) and for the medium risk stock (1-10 CFU/100 mL; 53% matching), possibly due to the smaller sample volume used (40 mL).
- When used as a presence/absence test, the Fluidion ALERT LAB correctly classified 93% of samples using a 1 CFU/100 mL cut-off, 95% of samples using a 10 CFU/100 mL cut-off, and 98% of samples using a 100 CFU/100 mL cut-off.
- In most cases where Fluidion ALERT LAB misclassified results, it showed a positive bias. It therefore overestimated, rather than underestimated, the risk.

Executive Summary

The primary concern regarding drinking water quality is that contamination of drinking water could lead to disease. A large number of pathogens can cause water-borne disease. The majority of these pathogens are fecal in origin, but it is not practical to test drinking water for all potential pathogens. Instead, measurement of fecal indicators is preferred. There is widespread agreement that *Escherichia coli* (*E. coli*) is the best currently available indicator of fecal contamination in drinking water.

A large number of test kits are available to quantify the presence of *E. coli* in water. The objective of this project has been to test and compare a range of kits against a certified reference method, which was chosen to be the IDEXX Quantitray 2000 method using Colilert medium. This report summarizes a set of laboratory assessments of different waters with different compositions and levels of contamination and presents the results of both the Fluidion ALERT LAB and the reference method.

The Fluidion ALERT LAB was compared to the reference method using cultivated *E. coli* in laboratory water with a phosphate-buffered saline matrix, as well as using wastewater treatment plant effluent diluted in five different sterilized natural waters (N1-N5). Results were interpreted graphically and through linear regression on both raw data and log-transformed data (see [Table 1](#)).

Table 1: Overview of the regression analysis for Fluidion ALERT LAB.

Water matrix	Number of paired samples	Maximum value	Slope (raw)	Intercept (raw)	Slope (log)	Intercept (log)	Spearman's r
Lab water	39	491	0.28	10.88	0.79	-0.03	0.961
N1	15	1168	2.66	5.11	1.01	0.41	0.924
N2	15	272	1.48	-0.81	0.90	0.22	0.909
N3	15	839	2.19	53.70	1.10	0.27	0.898
N4	15	442	1.84	3.72	0.88	0.48	0.907
N5	15	516	1.80	5.93	0.98	0.28	0.960

The Fluidion ALERT LAB was also assessed for false positives by using concentrated stocks of six non-target bacteria (*Aeromonas*, *Citrobacter*, *Enterobacter*, *Klebsiella*, *Pseudomonas aeruginosa* and *Serratia*); and for false negatives by using the same non-target bacteria spiked with low levels of *E. coli*. The Fluidion ALERT LAB did not report any false positive values in the absence of *E. coli* and was able to detect *E. coli* in the presence of each of the non-target bacteria.

Across the four test waters (excluding sterilized blanks) and five natural water matrices, a total of 60 paired samples were tested. In 87% of these, the semi-quantitative risk class matched the expected value. Matches were lowest for natural water N1 (75%) and for the medium risk stock (1-10 CFU/100 mL; 53% matching). The relative lack of concordance for the medium risk stock is to be expected as the Fluidion ALERT LAB uses a 40 mL volume for analysis, rather than 100 mL.

When used as a presence/absence test, the Fluidion ALERT LAB correctly classified 93% of samples using a 1 CFU/100 mL cut-off, 95% of samples using a 10 CFU/100 mL cut-off, and 98% of samples using a 100 CFU/100 mL cut-off.

Abbreviations

Colony Forming Unit	CFU
Defined Substrate Technology	DST
Ground water	GW
Lower Quantification Limit	LQL
Surface water	SW
Upper Quantification Limit	UQL

Contents

Abbreviations	3
Contents	4
1 Background information	5
2 Rapid Water Quality Testing project	5
3 Products	6
3.1 Trial Method: Fluidion ALERT LAB	6
3.2 Reference Method: IDEXX Quanti-Tray System	7
4 Test protocol and criteria	8
4.1 Phase 1	8
4.2 Phase 2	9
4.2.1 False Positives due to non-target bacteria	9
4.2.2 False negative due to competition	9
4.2.3 Natural waters	9
5 Results	11
5.1 Phase 1	11
5.2 Phase 2	12
5.2.1 False positive due to non-target bacteria.	12
5.2.2 False negatives due to competition	12
5.2.3 Natural waters	13
5.2.4 Natural waters spiked with effluent.	13
5.2.5 Statistical analysis Natural Waters.	14
5.3 Qualitative results	17
6 Appendix	18
6.1 Risk class matching	18
6.2 Traffic light assessment scheme.	21
6.2.1 False positive due to non-target bacteria.	22
6.2.2 False negatives due to competition	22
6.2.3 Natural waters	22
6.3 Manual	23

1 Background information

WHO and UNICEF both support national counterparts in monitoring and surveillance of drinking water quality in a variety of settings. In many countries where WHO and UNICEF work, logistical challenges mean that testing drinking water quality in laboratories is often not feasible, due to long distances and travel times required to transport samples. This has led to an interest in portable water quality testing kits, especially for measures of faecal contamination. Both WHO and UNICEF regularly procure portable water quality testing kits for national counterparts and share an interest in ensuring that the equipment procured can produce results that are reliable and match within reasonable margins the results from standard reference methods. In addition, both organizations wish to catalyse the continuous improvement of existing portable water quality testing products, and the development of innovative new products which might allow more efficient, accurate, or low-cost testing of drinking water quality in the field.

2 Rapid Water Quality Testing project

UNICEF, in collaboration with WHO, has developed a Rapid Water Quality Testing project to catalyse the continuous improvement of existing portable water quality testing products, and the development of innovative new products which might allow more efficient, accurate, or low-cost testing of drinking water quality in the field. The project has produced a Target Product Profile to describe the desired characteristics of a field test kit, and UNICEF has requested WHO to provide technical guidance on how to assess the performance of innovative products that result from the Rapid Water Quality Testing project.

There are a number of standards and methods used for measurement of microbiological quality of water, and many of the field test kits purport to follow these standards and methods. However, it can be difficult to conduct assessments with field kits out of a controlled laboratory environment, and some commercially available products, or innovative products recently developed, may in practice not meet all requirements.

In the absence of a clear procedure for assessing field test kits, the WHO Water, Sanitation and Hygiene team developed a template protocol for conducting such an assessment in a laboratory setting. This protocol has been reviewed by an independent technical advisory committee convened by WHO and UNICEF to support the Rapid Water Quality Testing project. The current protocol is focused on culture-based methods of measuring the faecal indicator bacterium *Escherichia coli* (*E. coli*).

The protocol consists of a first phase screening to determine if the assay under evaluation produces results comparable to the reference method over a range of *E. coli* concentrations, under highly controlled conditions. Assays that have passed Phase 1 assessments can proceed to the Phase of 2 of the assessment, which will examine the performance of the test under more challenging conditions (competition from non-target bacteria, use of different natural water matrices and wild *E. coli* strains, and variable temperature incubation if claimed by the manufacturer).

3 Products

3.1 Trial Method: Fluidion ALERT LAB

The Fluidion ALERT LAB is a remotely controllable, fully portable, and autonomous analyser for the measurement of *E. coli* and other bacteria. Suitable for source water and environmental monitoring at a field location, in a moving vehicle, or in a lab. It can perform up to six measurements at the same time using a 12V power source or battery.



Figure 1: FLUIDION ALTER LAB test kit.

Fluidion ALERT LAB is based on a modified real-time Defined Substrate Technology (DST) method, integrated into an automated instrument capable of performing bacterial quantification in the field. The device is capable of automating the incubation at 37°C, performing real-time multispectral optical analysis (absorbance/fluorescence), turbidity correction, signal analysis, and bacterial quantification of culturable *E. coli* with a time-to-result ranging between 2 and 12 hours. The technology tests 40 mL samples, and reportedly detects all bacteria present in solution, including those aggregated into clusters which might be interpreted as bacteria with standard methods.

Manufacturer information:

FLUIDION SAS

Bio&D, Et 1 C.C-ial Echat

94, Avenue du Général de Gaulle

94000 Créteil France

Phone: +33 182 390 290

Email: contact@fluidion.com

The Fluidion ALERT LABs with the following serial numbers were used during this validation study.

A52200005632

A52200005633

A52200005634

3.2 Reference Method: IDEXX Quanti-Tray System

The Colilert Test uses proprietary Defined Substrate Technology (DST) to simultaneously detect coliforms and *E. coli*. Two nutrient-indicators, ONPG and MUG, are the major sources of carbon in Colilert and can be metabolized by the coliform enzyme β -galactosidase and the *E. coli* enzyme β -glucuronidase, respectively.

Step 1

Add reagent to the sample.



Step 2

Pour into Quanti-Tray/2000 (counts from 1–2,419).



Step 3

Seal in Quanti-Tray Sealer and place in 35°C ± 0.5°C incubator for 24 hours.

(temperature requirement may be different per regulatory requirements in other countries)



Step 4

Yellow wells = total coliforms

Yellow/fluorescent wells = *E. coli*

Count positive wells and refer to MPN table



More information: <https://www.idexx.co.uk/en-gb/water/water-products-services/colilert/>

4 Test protocol and criteria

4.1 Phase 1

The first phase aimed to determine if the assay under evaluation produced results comparable to the reference method. This was done under highly controlled conditions over a range of *E. coli* concentrations.

A stock solution of a known lab strain of *E. coli* (ATCC 25922) with a concentration of approximately 1000 viable and culturable *E. coli* cells per 100 mL, was prepared (acceptable range: 300 - 3000 cells/100 mL). This was measured and confirmed using the IDEXX Quantitray method in a background of sterile phosphate buffered saline (pH 7.4 ± 0.2). This stock solution was then serially diluted using two-fold dilution with a sterile phosphate buffered saline, for details see [Table 2](#). The resulting stock solutions spanned a range of concentrations which were expected to yield positive results, ranging from zero to above most detection limits, with several critical range stock concentrations in between.

As a blank (A), a sample of stock solution 1 was autoclaved to eliminate any viable and culturable *E. coli*.

Table 2: Visual representation of the two-fold dilution, accounting for acceptable variance in the starting solution.

Stock	Approximate <i>E. coli</i> concentration, cells/100 mL		
	Lower acceptable limit	Target concentration	Upper acceptable limit
S1	300	1000	3000
S2	150	500	1500
S3	17	250	750
S4	38	125	375
S5	19	64	188
S6	9	32	94
S7	5	16	47
S8	2	8	23
S9	1	4	12
S10	0.6	2	6
S11	0.3	1	3
S12	0.1	0.5	1.5
A	0	0	0

The results of the assay under evaluation and the results of the reference method were plotted against each other using a log transformed linear regression of both datasets. Within a given stock, the triplicate samples from the assay under evaluation were “paired” with the triplicate analyses made with the reference method during sample processing (before the incubation period).

Samples below the minimum detection limit were fixed at 50% of the detection limit. Linear regression was made on the datapoints that were within the quantification range, or below the minimum detection limit, for both assays.

An assay proceeded to the Phase 2 assessment if the Spearman’s rank coefficient was at least 0.90, and if the blanks did not show positive results. It was originally intended that tests with a regression slope (before log transformation) significantly different from 1.0 would be excluded from Phase 2 assessment. However, a large number of trial assays had regression slopes significantly different from unity, so this condition was relaxed.

4.2 Phase 2

4.2.1 False Positives due to non-target bacteria

Some tests could potentially generate positive results in the absence of *E. coli* through the growth of non-target organisms. Cultures of six non-target bacteria (*Aeromonas*, *Citrobacter*, *Enterobacter*, *Klebsiella*, *Pseudomonas aeruginosa* and *Serratia*) that could potentially cause false positives, were made with a target concentration of 100,000,000 viable and culturable cells/100 mL (acceptable range: 30,000,000 – 300,000,000 cells/100 mL). These cultures were tested using the trial assay without any addition of *E. coli*. Any positive results were considered a false positive. Single tests instead of triplicates were done, and the reference method was not challenged with the non-target organisms.

4.2.2 False negative due to competition

The same six cultures of non-target organisms were mixed 1:100 with *E. coli* Stock 1, resulting in an approximate concentration of 30 CFU/100 mL *E. coli* and 30,000 CFU/100 mL of the non-target organism. The resulting stock was tested using the trial kit. Any negative results were considered to indicate that in the presence of competing bacteria, *E. coli* might not be detected by the trial method.

As for the False Positive experiments, the reference method was not tested and only single tests instead of triplicates were done.

4.2.3 Natural waters

The water matrix, as well as the strain of *E. coli* used, may affect the performance of the trial method. To assess this possibility, five different natural waters were selected. These included at least two surface water (SW) and two groundwater (GW) sources. Full list of requirements for the natural waters can be found in [Table 3](#).

Table 3: Criteria for the natural waters.

Natural water	Source	Turbidity	pH	Alkalinity
N1	GW or SW	> 10	Any	At least one of the waters should have a low <50 mg/L CaCO ₃
N2	GW or SW	< 10	< 6.5	
N3	GW or SW	< 10	> 8.0	
N4	GW or SW	Any	6.5 – 8.0	
N5	GW or SW	Any	Any	

The natural waters were sterilised and then spiked with unfiltered effluent from a wastewater treatment plant to reach a target concentration of 300 *E. coli* per 100 mL (acceptable range: 100 – 1000 cells/100 mL). Pre-testing of the effluent was required to determine the concentration in order to properly dilute it into the natural waters. The stock solutions of effluent in natural water were serially diluted using ten-fold dilutions with the sterilised natural waters three times. The resulting stock solutions spanned a range of concentrations which would be expected to yield at least one stock in each of the risk classes listed below in [Table 4](#). The blank (A) was made by autoclaving the natural waters.

Table 4: Ten-fold dilution of effluent stock solution in sterilised waste water, accounting for the acceptable variance in starting solution.

Stock	Risk class	Approximate <i>E. coli</i> concentration, cells/100 mL		
		Lower acceptable limit	Target concentration	Upper acceptable limit
N*S1	Very high	100	300	1000
N*S2	High	10	30	100
N*S3	Medium	1	3	10
N*S4	Low	0.1	0.3	1
N*A	Not applicable	0	0	0

All natural water stocks were tested in triplicate with the trial method, using three different sets of equipment per triplicate: 5 water stocks (N1-5) * 5 dilution stocks (N*S1-A) * 3 replicates using different equipment, for a total of 75 analyses in all (60 stocks and 15 blanks). The same was done for the reference method.

Samples below the minimum detection limit were fixed at 50% of the detection limit. Linear regression was made on the datapoints that were within the quantification range, or below the detection limit, for both assays. Statistical tests were made as in Phase 1.

5 Results

5.1 Phase 1


Tests were performed by one technician. The stock dilutions were made the day of testing. For the FLUIDION ALTER LABs it was not possible to analyse all the stocks on the same day due to time restraints. This means that the blanks were measured the day before the test. The next day, the stock solutions made with *E. coli* suspension, and stock 1-6 were analysed. The day after (within 24 hours), stocks 7-12 were tested. This is reflected in the results of the reference method.

Results were compared to the reference method over a wide range of *E. coli* concentrations, under highly controlled conditions (see Table 5).

Table 5: Results of the CFU testing using the reference method and trial method over multiple dilutions and different sets of equipment.

Stock	Reference method (CFU/100 mL)			Trial method (CFU/100 mL)		
	1	2	3	A52200005632	A52200005633	A52200005634
S1	1553	1120	1986	491	491	425
S2	727	727	816	217	251	247
S3	291	461	387	217	143	111
S4	152	162	155	85	56	83
S5	114	82	86	49	43	25
S6	29	46	35	25	22	13
S7	47	64	81	4	7	6
S8	22	17	23	6	3	6
S9	7.5	12.1	9.8	< 2.5	2.5	4
S10	3.1	5.2	4.1	2.5	< 2.5	4
S11	1	2	2	< 2.5	2.5	< 2.5
S12	1	1	2	< 2.5	< 2.5	3
Blank	< 1.0	< 1.0	< 1.0	< 2.5	< 2.5	< 2.5

Cultivated *E.coli*, Fluidion



Incubation temperature: Built in. Incubation duration: Built in. 39 paired samples.

Figure 1: Statistical analysis Cultivated *E.coli*.

The Spearman's rank coefficient = 0.961 and meets the criterion. The slope of the un-transformed regression was significantly less than unity. However, because the same was true of many other trial assays, and because the Fluidion ALERT LAB is very different from other commercially available assays, WHO and UNICEF agreed to proceed to Phase 2 assessments with the Fluidion ALERT LAB.

5.2 Phase 2

5.2.1 False positive due to non-target bacteria.

The results of the false positives test can be found in [Table 6](#) below. The full list of numerical results can be found in Appendix 6.2.1.

Table 6: Results of the false positives test

Non-target bacteria (100,000,000 CFU/100 mL)	Target bacteria (30 CFU/100 mL)	Test results
<i>Aeromonas</i>		negative
<i>Citrobacter</i>		negative
<i>Enterobacter</i>		negative
<i>Klebsiella</i>		negative
<i>Pseudomonas</i>		negative
	<i>E. coli</i> *	positive

* *E. coli* has been analysed as a positive control to ensure growth conditions.

5.2.2 False negatives due to competition

The results of the false positives test can be found below in [Table 7](#). The full list of numerical results can be found in Appendix 6.2.2.

Table 7: Results of the false negatives test.

Non-target bacteria (30,000 CFU/100 mL)	Target bacteria (30 CFU/100 mL)	Test results
<i>Aeromonas</i>	<i>E. coli</i>	positive
<i>Citrobacter</i>	<i>E. coli</i>	positive
<i>Enterobacter</i>	<i>E. coli</i>	positive
<i>Klebsiella</i>	<i>E. coli</i>	positive
<i>Pseudomonas</i>	<i>E. coli</i>	positive
	<i>E. coli</i>	positive

5.2.3 Natural waters

pH, turbidity, and alkalinity of all natural water samples were tested and matched with the criteria from Table 3. Since autoclaving the water samples caused changes in the pH and turbidity, some natural waters were sterilized by filtration through 0.22 µm filters in order to meet the target specifications (see below in Table 8).

Table 8: Selection of the natural water samples and their required and tested specifications.

Waters	Sample point coding	Matrix	Sterilization	Specifications	Required	Tested
N1	Supply channel after Bethune polder pumping station	SW	Autoclave	pH	any	8.4
				Turbidity (FTU)	> 10	89
				Alkalinity (mg/L)	any	210
N2	Pumping station Archemberg joint raw groundwater	GW	Filtration 0.22 µm	pH	< 6.5	6.2
				Turbidity (FTU)	< 10	< 0.1
				Alkalinity (mg/L)	any	18
N3	Surface water intake point on the Petrusplaat	SW	Autoclave	pH	> 8	8.3
				Turbidity (FTU)	< 10	3.4
				Alkalinity (mg/L)	any	50
N4	Pumping station Nijmegen joint raw ground water	GW	Filtration 0.22 µm	pH	6.5 - 8.0	7.5
				Turbidity (FTU)	any	< 0.1
				Alkalinity (mg/L)	any	55
N5	Pumping station Vessum joint raw ground water	GW	Filtration 0.22 µm	pH	any	6.6
				Turbidity (FTU)	any	5.7
				Alkalinity (mg/L)	< 50	22

5.2.4 Natural waters spiked with effluent.

In Table 9, the results for the measurement of colony forming units using both the reference and the trial method can be found. This was done for all the natural water sample with different effluent concentrations. A total of 15 paired samples were analysed for each natural water, for a grand total of 75 paired samples, including 15 blanks. No *E. coli* was detected in any of the blank samples, using either the trial or reference method.


Table 9: Results in CFU/100 mL of the natural waters spiked with effluent from the wastewater treatment plant for both the reference and trial method.

Stock	Replicate	N1		N2		N3		N4		N5	
		Ref	Trial								
S1	1	344.1	718	191.8	233	109.5	839	135.4	272	204.6	516
	2	325.5	999	167	272	435.2	718	195.6	442	290.9	516
	3	410.6	1168	119.8	228	193.5	839	214.3	318	325.5	516
S2	1	24.6	46	18.9	13	22.3	74	13.5	53	28.8	45
	2	21.6	101	18.7	18	27.9	74	26	62	23.8	53
	3	21.8	46	23.1	28	26.5	53	30.9	13	21.3	62
S3	1	5.2	15	3.1	< 2.5	2	2.5	1	24	4.1	4
	2	3.1	73	3.1	< 2.5	1	2.5	3.1	9	2	< 2.5
	3	4.1	2.5	3	4	1	< 2.5	< 1	2.5	3.1	2.5
S4	1	< 1	< 2.5	< 1	< 2.5	1	< 2.5	< 1	< 2.5	< 1	< 2.5
	2	< 1	< 2.5	< 1	< 2.5	2	< 2.5	< 1	< 2.5	< 1	< 2.5
	3	1	< 2.5	< 1	< 2.5	< 1	< 2.5	1	< 2.5	< 1	< 2.5
A	1	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5
	2	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5
	3	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5	< 1	< 2.5

5.2.5 Statistical analysis Natural Waters.

Graphical interpretation and overview of results on both raw data and log-transformed data for all five natural water matrices can be found below in [Figure 2](#) to [Figure 6](#).


Matrix N1, Fluidion



Incubation temperature: Built in. Incubation duration: Built in. 15 paired samples.

Figure 2: Statistical analysis Natural Matrix N1.

Matrix N2, Fluidion



Incubation temperature: Built in. Incubation duration: Built in. 15 paired samples.

Figure 3: Statistical analysis Natural Matrix N2.

Matrix N3, Fluidion




Figure 4: Statistical analysis Natural Matrix N3.

Matrix N4, Fluidion





Figure 5: Statistical analysis Natural Matrix N4.

Matrix N5, Fluidion



Incubation temperature: Built in. Incubation duration: Built in. 15 paired samples.

Figure 6: Statistical analysis Natural Matrix N5.

Interpretation and overview of results through linear regression on both raw data and log-transformed data is summarised below in Table 10.

Table 10: Overview of the regression analysis for Fluidion ALERT LAB.

Water matrix	Number of paired samples	Maximum value	Slope (raw data)	Intercept (raw data)	Slope (log trans)	Intercept (log trans)	Spearman's r
Lab water	39	491	0.28	10.88	0.79	-0.03	0.961
N1	15	1168	2.66	5.11	1.01	0.41	0.924
N2	15	272	1.48	-0.81	0.90	0.22	0.909
N3	15	839	2.19	53.70	1.10	0.27	0.898
N4	15	442	1.84	3.72	0.88	0.48	0.907
N5	15	516	1.80	5.93	0.98	0.28	0.960

The trial method was also assessed using the semi-quantitative risk classes defined in Table 4. An analysis was considered to correctly match the risk class if stock 1 yielded a result above 100 CFU/100 mL, if stock 2 yielded a result of at least 11 and no more than 100 CFU/100 mL, if stock 3 yielded a result of at least 1 and no more than 10 CFU/100 mL, and if stock 4 had either no detectable *E. coli* or a maximum of 1 CFU/100 mL. Detailed tables for each natural water matrix are shown in 6.1, and Table 11 below presents a summary, showing the overall, 87% of analyses gave the correct risk class.

Table 11: Results matching expected risk class. (% results in risk class)

Test Water	Risk Class	Water Matrix					Average
		N1	N2	N3	N4	N5	
S1	>100 CFU/100 mL (very high risk)	100%	100%	100%	100%	100%	100%
S2	11-100 CFU/100 mL (high risk)	67%	100%	100%	100%	100%	93%
S3	1-10 CFU/100 mL (medium risk)	33%	33%	67%	67%	67%	53%
S4	<=1 CFU/100 mL* (low risk)	100%	100%	100%	100%	100%	100%
Grand Average							87%

Finally, the utility of the test to produce dichotomous presence/absence results was assessed at different thresholds. With a threshold of 1 CFU/100 mL, 98% of tests were correctly classified. With thresholds of 10 and 100 CFU/100mL, the proportion of tests correctly classified were 95% and 98%, respectively (see Table 12).

Table 12: Summary of presence/absence results for Fluidion ALERT LAB.

Water matrix	Presence/absence cut-off		
	1 CFU/100 mL	10 CFU/100 mL	100 CFU/100 mL
N1	100%	83%	92%
N2	83%	100%	100%
N3	92%	100%	100%
N4	100%	92%	100%
N5	92%	100%	100%
All	93%	95%	98%

5.3 Qualitative results

Lastly, a qualitative assessment of the FLUIDION test kits was made with reference to categories ranging from the ease of use to the safety of the user and environment. Summary of these results can be found below in Table 13.

Table 13: Scoring of the user friendliness of the FLUIDION test kits.

Subjects	Assessment	Explanation
User manual		
Contact with a remote server	Variable	If you use google to go to the page of FLUIDION it will be better. It is important to first bring the device into contact with the site before taking the sample.
Execution test	Easy	
Interpretation results	Easy	Results are clearly visible, per sample, visible on the site of FLUIDION.
Contamination risk to:	Sample	Low
	User	Low
Dispose of materials with a high concentration of <i>E. coli</i>	Easy	The manual recommends disposing of used samples following the guidelines of the country of operation, or adding liquid bleach to disinfect the samples.

6 Appendix

6.1 Risk class matching

Table 14: Risk class matching expected risk class, Natural Water N1.

Test water	Risk class	CFU/mL	Risk class difference	Correct risk class	
				Single test	Triplicates
S1	>100 CFU/100 mL (very high risk)	718	0	100%	100%
		999	0	100%	
		1168	0	100%	
S2	11-100 CFU/100 mL (high risk)	46	0	100%	67%
		101	1	0%	
		46	0	100%	
S3	1-10 CFU/100 mL (medium risk)	15	1	0%	33%
		73	1	0%	
		2.5	0	100%	
S4	<=1 CFU/100 mL (low risk)	< 2.5	0	100%	100%
		< 2.5	0	100%	
		< 2.5	0	100%	
Average			0.25	75%	
Presence/Absence (1 CFU cut-off)				100%	
Presence/Absence (10 CFU cut-off)				83%	
Presence/Absence (100 CFU cut-off)				92%	

Table 15: Risk class matching expected risk class, Natural Water N2.

Test water	Risk class	CFU/mL	Risk class difference	Correct risk class	
				Single test	Triplicates
S1	>100 CFU/100 mL (very high risk)	233	0	100%	100%
		272	0	100%	
		228	0	100%	
S2	11-100 CFU/100 mL (high risk)	13	0	100%	100%
		18	0	100%	
		28	0	100%	
S3	1-10 CFU/100 mL (medium risk)	< 2.5	-1	0%	33%
		< 2.5	-1	0%	
		4	0	100%	
S4	<=1 CFU/100 mL (low risk)	< 2.5	0	100%	100%
		< 2.5	0	100%	
		< 2.5	0	100%	
Average			0.17	83%	
Presence/Absence (1 CFU cut-off)				83%	
Presence/Absence (10 CFU cut-off)				100%	
Presence/Absence (100 CFU cut-off)				100%	

Table 16: Risk class matching expected risk class, Natural Water N3.

Test water	Risk class	CFU/mL	Risk class difference	Correct risk class	
				Single test	Triplicates
S1	>100 CFU/100 mL (very high risk)	839	0	100%	100%
		718	0	100%	
		839	0	100%	
S2	11-100 CFU/100 mL (high risk)	74	0	100%	100%
		74	0	100%	
		53	0	100%	
S3	1-10 CFU/100 mL (medium risk)	2.5	0	100%	67%
		2.5	0	100%	
		< 2.5	-1	0%	
S4	<=1 CFU/100 mL (low risk)	< 2.5	0	100%	100%
		< 2.5	0	100%	
		< 2.5	0	100%	
Average			0.08	92%	
Presence/Absence (1 CFU cut-off)				92%	
Presence/Absence (10 CFU cut-off)				100%	
Presence/Absence (100 CFU cut-off)				100%	

Table 17: Risk class matching expected risk class, Natural Water N4.

Test water	Risk class	CFU/mL	Risk class difference	Correct risk class	
				Single test	Triplicates
S1	>100 CFU/100 mL (very high risk)	272	0	100%	100%
		442	0	100%	
		318	0	100%	
S2	11-100 CFU/100 mL (high risk)	53	0	100%	100%
		62	0	100%	
		13	0	100%	
S3	1-10 CFU/100 mL (medium risk)	24	1	0%	67%
		9	0	100%	
		2.5	0	100%	
S4	<=1 CFU/100 mL (low risk)	< 2.5	0	100%	100%
		< 2.5	0	100%	
		< 2.5	0	100%	
Average			0.08	92%	
Presence/Absence (1 CFU cut-off)				100%	
Presence/Absence (10 CFU cut-off)				92%	
Presence/Absence (100 CFU cut-off)				100%	

Table 18: Risk class matching expected risk class, Natural Water N5.

Test water	Risk class	CFU/mL	Risk class difference	Correct risk class	
				Single test	Triplicates
S1	>100 CFU/100 mL (very high risk)	516	0	100%	100%
		516	0	100%	
		516	0	100%	
S2	11-100 CFU/100 mL (high risk)	45	0	100%	100%
		53	0	100%	
		62	0	100%	
S3	1-10 CFU/100 mL (medium risk)	4	0	100%	67%
		< 2.5	-1	0%	
		2.5	0	100%	
S4	<=1 CFU/100 mL (low risk)	< 2.5	0	100%	100%
		< 2.5	0	100%	
		< 2.5	0	100%	
Average			0.08	92%	
Presence/Absence (1 CFU cut-off)				92%	
Presence/Absence (10 CFU cut-off)				100%	
Presence/Absence (100 CFU cut-off)				100%	

6.2 Traffic light assessment scheme.

In order to assist with the interpretation of the Phase 2 results, the following 'traffic light' assessment scheme is used, in which results are considered to be 'green' if the results meet the statements listed in the kit's manual, 'yellow' if there is some disparity between results and the expected results, or there is a potential risk of infection to the user, and 'red' if the results deviate significantly from the expected results. The detailed assessment scheme is described below.

Results do not meet the guidelines listed in the kit's manual.



False positives:

Two or more tests are positive

False negatives:

Two or more tests are negative


Incubation temperature:

A score is given to each temperature and the score deviates by a factor of more than 2.

Natural waters:

The results match the expected risk class less than 50% of the time in at least one natural water matrices, or less than 80% of the time in at least three natural water matrices

Disparity between results and the kit's guidelines compared to the potential risk to the user.



False positives:

If only one test is positive. Risk of infection to the user is minimal.

False negatives:

One test is negative


Incubation temperature:

A score is given to each temperature. If the score does not deviate by a factor of more than 2, the results stay in the same risk class.

Natural waters:

The results match the expected risk class at least 50% of the time in all five natural water matrices, and at least 80% of the time in at least three natural water matrices.

Results meet the guidelines listed in the kit's manual.



False positives:

None of the tests are positive.

False negative:

All the tests are positive.

Incubation temperature:

Incubation results matches the temperature range in the kit's manual.

Natural waters:

The results match the expected risk class at least 80% of the time in all five natural water matrices, and at least 90% of the time in at least three natural water matrices.

6.2.1 False positive due to non-target bacteria.

Table 19: Results of the false positives test

Non-target bacteria (1*10 ⁸ CFU/100 mL)	Quantitative test results (CFU/100 mL)			
	A5220005632	A5220005633	A5220005634	
<i>Aeromonas</i>	< 2.5	< 2.5	< 2.5	
<i>Citrobacter</i>	< 2.5	< 2.5	< 2.5	
<i>Enterobacter</i>	< 2.5	< 2.5	< 2.5	
<i>Klebsiella</i>	< 2.5	< 2.5	< 2.5	
<i>Pseudomonas</i>	< 2.5	< 2.5	< 2.5	
<i>E. coli</i> *	1.35*10 ⁷	1.37E*10 ⁷	1.37E*10 ⁷	

* *E. coli* has been analysed as a positive control to ensure growth conditions.

6.2.2 False negatives due to competition


Table 20: Results of the false negatives test.

Non-target bacteria (30,000 CFU/100 mL)	Target bacteria (30 CFU/100 mL)	Quantitative test results (CFU/100 mL)			
		A5220005632	A5220005633	A5220005634	
<i>Aeromonas</i>	<i>E. coli</i>	9	11	11	
<i>Citrobacter</i>	<i>E. coli</i>	39	50	59	
<i>Enterobacter</i> *	<i>E. coli</i>	324	85	549	
<i>Klebsiella</i> *	<i>E. coli</i>	247	98	191	
<i>Pseudomonas</i>	<i>E. coli</i>	38	29	43	
	<i>E. coli</i>	33	25	25	

*Results are much higher. The number of *E. coli* added are approximately 30 CFU/100 mL.

6.2.3 Natural waters

Table 21: Results matching expected risk class, by water matrix.

Test Water	Water Matrix					
	N1	N2	N3	N4	N5	
S1	100%	100%	100%	100%	100%	
S2	67%	100%	100%	100%	100%	
S3	33%	33%	67%	67%	67%	
S4	100%	100%	100%	100%	100%	
Average	75%	83%	92%	92%	92%	
Grand Average	87%					

Note that this yellow light is driven by three non-concordant samples with water matrix N1 (table 14). One of these results was extremely close to the threshold: if the second replicate for test water S2 had read "100" rather than "101" it would have been within the target risk class, and the overall average for the N1 water matrix would have been 83%, resulting in a green light assessment. The yellow light rating should therefore be considered in context.

6.3 Manual



ALERT LAB


Manual V3.46

Table of Contents

System Overview	3
Operation and Maintenance	7
Web Interface	18
Command Portal	21
Data Management Interface	25

System Overview

System Configuration



System Configuration

Handle



Battery holder



Specifications Overview

TECHNICAL SPECIFICATIONS

Dimensions	<i>26x24x22 cm</i>	Total measurements	<i>6</i>
Weight	<i>3 kg</i>	Response time	<i>2h-14 h</i>
Measurement trigger	<i>On-demand</i>	Communication	<i>GSM, USB, secure web interface</i>
Measurement range	<i>4 CFU - 1x10⁶ CFU/100 mL</i>	Antenna	<i>Internal</i>
Materials	<i>PMMA, PP, Acetal, SST 316L, glass</i>	Alert notifications	<i>email (optional)</i>


Operation and Maintenance



Safety Precautions



-  *Wear gloves to protect yourself and avoid contamination of the sample*
-  *Wear safety glasses whenever handling broken glass. Decontaminate hands / gloves prior to touching eyes.*
-  *All waste produced by microbiology methods must be discarded following the local biohazard waste management regulation.*
-  *Never short the power connector or allow water to enter the system.*
-  *Read and follow the instructions of all MSDS sheets for the reagents used*



System Power



Use only the power supply adapter
or Li-ion battery provided!

Connect the power adapter located right side of the system.

Turn ON: Set power switch to ON position

Turn OFF: Set power switch to OFF position

Battery holder



Battery Box



Rail
(Battery holder)

Insert the battery box inside the Rail located behind the system

Please make sure the battery box is in the correct orientation

Plug the power cable into the Battery box

Bioreagent filling – fresh water



*Ensure proper microbiological
decontamination of the vials
prior to usage*


Inject 1 mL of ALERT Bioreagent E.coli or Entero in each sterile vial using syringe provided.

Add 25 mL of sample.

Ensure cap has rubber gasket correctly installed (only for Pyrex round bottom vials) and close the vial.

Stir gently to homogenize the solution.

Bioreagent filling – sea water



Ensure proper microbiological decontamination of the vials prior to usage

Inject 1 mL of ALERT Bioreagent E.coli or Entero in each sterile vial using syringe provided.

Add 27 mL of sterile water.

Add 9 mL of sample.

Ensure cap has rubber gasket correctly installed (only for Pyrex round bottom vials) and close the vial.

Stir gently to homogenize the solution.

Bioreagent filling – powder reagent

Step 1: Transfer 40mL of sample



Step 2: Add one "snap-pack" of powder bioreagent to sample



Step 3: Mix



Step 4a: Insert in device and launch the analysis



- Add 40ml of sample to single-use plastic vial
- Add one pack of powder bioreagent to sample
- Mix until dissolved
- Insert in device and start the analysis

LED indicator status

When the system is powered on, LEDs indicate the system operation status.



Blue GSM Light:

- Blinking rapidly (once / second) → Searching for Network
- Blinking slowly (once / 3 seconds) → Network connection established

Orange GPS Light:

- Blinking rapidly (once / second) → GPS activated

Red Measurement Status Light

- LED ON for a numbered cell → Measurement on going



Best Practice:

Send “**PING**” command to the system to make sure communication is working before starting a measurement (see “**Command Portal**” section).

Operating Procedure



To operate the system, you must be an authorized user and Login through the Secure Fluidion website (see “*Command Portal*” section)

Note: After setting the power switch to the ON position, the system will take 1-2 minutes to establish connection to the mobile network.

1 - Login to the Command Portal and select the device you wish to control using the dropdown menu.

2 - Click the « **PING** » button to get the device status and wait for the device response.

Verify if the date, incubation time and other parameters are correct.

3 - To start an individual measurement, insert the sample vial in the cell X and click the « **START MEASUREMENT** » button. Select the cell(s) to be activated and click the « **CONFIRM** » button.

(X=number between 1 and 6)

(An optional text label can be added to each measurement for easy identification)

*For advanced detail see the « **Command Portal** » section.*

Reusable glass vial cleaning




Ensure working area is clean to avoid any external contamination.

Never reuse single-use plastic vials.

Disinfect vials with bleach solution (**10%**), clean using brush and autoclave following regular microbiological cleaning procedures.

Clean white plastic caps with bleach solution. Do not autoclave.

Dispose of the sample according to local biological waste regulations.



Sample disposal – single-use plastic vials

The contents of the plastic tube should be disposed properly as per the guidelines in the country of operation. Typically, this liquid solution is classified as hazardous biological laboratory waste and needs to be treated accordingly.

In the event that certified waste disposal services are not available in the place of operation, a small amount (minimum 4mL) of commercially available freshly-prepared liquid bleach solution (typically 3-6% sodium hypochlorite), should be added to each plastic vial which contains bio-waste solution. The solution with bleach should be allowed to stand for a minimum of 15 minutes contact time, to ensure that all bacteria within the solution have been deactivated. This solution should then be safe to dispose; example: by flushing down a toilet.


The plastic vials should never be re-used.



Web Interface



Data Account

www.secure-fluidion.com



- Real-time data access
- Device management
- User management
- Maintenance

Data Account



Command Portal

Command Portal (Quick Commands)

Data Management Interface ← **Data Management interface button**

System: A5221005635 (A5221005635)

System Status

Last sync: 2022-02-03 14:19:35

Measurement: E. Coli/Total Coliform Fresh Water 1

Network: Orange

Last communication: 2022-02-03 14:37:58

Recent measurements

2022-02-02 17:05:27 (Europe/Paris), Cell: 3.
- Label: .
- E. Coli: ≤4 /100mL(Finished, 14.0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14.0 hours)
- Get report

2022-02-02 17:05:27 (Europe/Paris), Cell: 4.
- Label: .
- E. Coli: ≤4 /100mL(Finished, 14.0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14.0 hours)
- Get report

2022-02-02 17:05:27 (Europe/Paris), Cell: 5.
- Label: .

Quick Commands

Ping **Start Measurement**

Advanced Commands ▾

Total: 71 records Page: 1/3 << < > >>

Command	Operator	Created at	Status
PING S/N: A5221005635 DATETIME: 2022-02-02 10:02:57 (Europe/Paris) BAT: 12.71 SIG: 29 INC TIME: 14 MEASURE: E.COLI ONGOING: GPS: ON VS.46: Received at: 2022-02-02 17:00:15	Omar Bach Rais	2022-02-02 17:02:51	Finished
START MEASUREMENT 1,2,3,4,5,6 2022/02/01 15:52:45 MEASUREMENT 1,2,3,4,5,6 STARTED; Received at: 2022-02-01 15:54:00	Omar Bach Rais	2022-02-01 15:51:48	Finished

This will send a **A5221005635** command to system:

Please select the cells that you wish to Start Measurement.

Cell #	Measurement Label (Optional)
1 <input checked="" type="checkbox"/>	Sample 1
2 <input checked="" type="checkbox"/>	Sample 2
3 <input checked="" type="checkbox"/>	Sample 3
4 <input type="checkbox"/>	
5 <input type="checkbox"/>	
6 <input type="checkbox"/>	

Select all

Confirm **Cancel**

Command Portal (Advanced Commands)

Data Management Interface

System: A5221005635 (A5221005635)

System Status

1 **Orange** 2 **Orange** 3 **Orange**
4 **Ready** 5 **Ready** 6 **Ready**

Last sync:
2022-02-03 14:19:28

Measurement:
E. Coli/Total Coliform Fresh Water 1

Network:
Orange

Last communication:
2022-02-03 14:17:58

Recent measurements

2022-02-02 17:05:27 (Europe/Paris), Cell: 3.
- Label: ...
- E. Coli: ≤4 /100mL(Finished, 14,0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14,0 hours)
- Get report

2022-02-02 17:05:27 (Europe/Paris), Cell: 4.
- Label: ...
- E. Coli: ≤4 /100mL(Finished, 14,0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14,0 hours)
- Get report

2022-02-02 17:05:27 (Europe/Paris), Cell: 5.
- Label: ...

Quick Commands

Ping Start Measurement

Advanced Commands ▾

Synchronization Stop measurements Download data Get result Get GPS GPS On GPS Off **Set measurement type**

Type in the command:
Your message Send

Attention: Do NOT upload new firmware and new certificate at the same time. Please make sure the previous upload are finished.

Select the firmware file:
Choose File No file chosen
Send

Select the certificate file:
Choose File No file chosen
Send

Command	Operator	Created at	Status
PING S/N: A5221005635 DATETIME: 2022-02-02 16:02:57 (Europe/Paris) BAW: 12.71 SIG: 29 INC TIME: 14 MEASURE: E.COLI ONGOING: GPS: ON VS.4R: <small>Received at: 2022-02-02 17:04:10</small>	Omar Bach Raie	2022-02-02 17:02:51	Finished
START MEASUREMENT 1,2,3,4,5,4 2022/02/01 15:52:45 MEASUREMENT 1,2,3,4,5,6 STARTED; <small>Received at: 2022-02-01 17:04:00</small>	Omar Bach Raie	2022-02-01 15:51:48	Finished

This will send a **A5221005635** command to system:

Please select the cells that you wish to Stop Measurement.

Cell #

1
2
3
4
5
6
Select all

Confirm Cancel

Measurement Type:

Please select the measurement type:

E. Coli/Total Coliform Fresh Water 1 ▾


Confirm Cancel

Command Portal (Command Responses)

Data Management Interface

System:

System Status



fluidion

Last sync:
2022-02-03 14:19:35

Measurement:
E. Coli/Total Coliform Fresh Water 1

Network:
Orange

Last communication:
2022-02-03 14:37:58

Recent measurements

2022-02-02 17:05:27 (Europe/Paris), Cell: 3.
- Label: .
- E. Coli: ≤4 /100mL(Finished, 14.0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14.0 hours)
- [Get report](#)

2022-02-02 17:05:27 (Europe/Paris), Cell: 4.
- Label: .
- E. Coli: ≤4 /100mL(Finished, 14.0 hours)
- Total Coliform: ≤4 /100mL(Finished, 14.0 hours)
- [Get report](#)

2022-02-02 17:05:27 (Europe/Paris), Cell: 5.
- Label: .

set measurement ecoli

MEASUREMENT SET: E.COLI;
Received at: 18 Jun, 2020 15:36:56

STOP MEASUREMENT 1,2,3,4,5,6

MEASUREMENT 1,2,3,4,5,6 HAS BEEN STOPPED;
Received at: 18 Jun, 2020 14:22:25

download data

DATA DOWNLOAD SUCCESSFUL;
Received at: 2 Jun, 2020 14:25:04

Quick Commands

[Ping](#) [Start Measurement](#)

Advanced Commands ▾

Total: 71 records Page: 1/3 << < > >>

Command	Operator	Created at	Status
PING	Omar Bach Rais	2022-02-02 17:02:51	Finished
<p>S/N: A5221005635 DATETIME: 2022-02-02 10:02:57 (Europe/Paris) BAT: 12.71 SIG: 29 INC TIME: 14 MEASURE: E.COLI ONGOING: GPS: ON VS.46: Received at: 2022-02-02 17:00:15</p>			
START MEASUREMENT 1,2,3,4,5,6	Omar Bach Rais		
<p>2022/02/01 15:52:45 MEASUREMENT 1,2,3,4,5,6 STARTED; Received at: 2022-02-01 15:54:00</p>			

GET GPS

LAT: 48.796889
 LONG: 2.448735
 SAT VIEWED: 13
 SAT USED: 4
 SIGNAL[0-55]: 27;
 Received at: 8 Jun, 2020 11:56:30
[Google Maps](#)

GET RESULT

Cell 1 Label:
 Started: 2020-06-18 14:02:37 (3.2h)
 Calibration: E. Coli/Total Coliform Fresh Water 1
 Total Coliform: ≤6.58×10⁵ /100mL (In progress)
 E. Coli: ≤5.48×10⁶ /100mL (In progress)

Data Management Interface

Data Management Interface (Front Page)

The screenshot displays the fluidion Data Management Interface (Front Page). The interface includes a navigation bar with tabs for 'Devices', 'Data History', and 'User Center'. A sidebar on the left lists 'My devices' including 'ALERT System', 'AR-ALERT System', 'ALERT LAB', 'ALERT System V2', and 'ALERT V2 - Iowa Exch I...'. The main content area features a map, a 'System general info' table, 'Bottle status', 'Schedules', and 'Operations' sections. A 'Command Portal' button is highlighted in red. Below the main content, a navigation bar contains 'Infer sensor data', 'Measurement history', 'Commands history', and 'System maintenance'. The 'Measurement history' section shows a graph of 'Measurements in system' with 'Absorbance' and 'Fluorescence' data series. The graph shows a sharp increase in absorbance and fluorescence starting around 2021-12-22. The 'Measurement Results' table shows 'E. coli concentration: 608 /100mL' and 'Total coliform concentration: 5111 /100mL'. The 'Command Portal' button is highlighted in red. The 'Page Tabs' are also highlighted in red.

System general info

VN	ALCJ1008812
Label	ALERT V2 - Iowa Exch I...
Battery	13.19V
DSM signal	18 (DATA)
Last update	2021-12-23 09:18:21
Status	Activated (Deep Sleep mode: OFF)
Measurement type	E. Coli/Total Coliform
Default calibration	E. coli / TC ALERT System V2 Freshwater - Beta 2.0
Firmware version	V1.21
GPS status	ON
Last GPS update	2021-06-09 22:40:58
Location	48.819479,2.292961

Command Portal button

Page Tabs

Data Interface (Ongoing Measurements)



Sample time and location

Ongoing measurements

Current results

Optical data

Generate report button

Export data button

System advanced data

Data Interface (Measurement History)

The screenshot shows a web interface with a navigation bar at the top containing 'Inline sensor data', 'Measurement history' (highlighted with a red box), 'Commands history', and 'System maintenance'. Below the navigation bar is a 'Filters' section on the left, which includes a dropdown menu for 'All records', a checked checkbox for 'Include historic measurements', and several unchecked checkboxes for 'Start date', 'End date', and 'GPS Range'. The main area displays a table of measurement records. Each record includes an ID, Operator, Bottle number, Label, Tag, Result, Date, and Location. The 'Result' field contains 'E. Coli: ≤ 4 /100ml.' and 'Total Coliform: ≤ 4 /100ml.'. The 'Date' is '2022-02-02 17:05:27' and the 'Location' is '48.795621,2.451875'. Each record has two buttons at the bottom: 'show data graph' and 'generate report'. A red box highlights the 'generate report' button for record ID 169847. A red arrow points from the 'Measurement history search bar' label to the 'Search' button in the filters section. Another red arrow points from the 'Historical measurement' label to the details of record ID 169847. A third red arrow points from the 'Generate report button' label to the 'generate report' button for record ID 169852. A fourth red arrow points from the 'Show data graph button' label to the 'show data graph' button for record ID 169852.

ID	Operator	Information	Status
169850	Omar Bach Rais	Bottle: 6 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished
169849	Omar Bach Rais	Bottle: 5 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished
169848	Omar Bach Rais	Bottle: 4 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished
169847	Omar Bach Rais	Bottle: 3 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished
169852	Omar Bach Rais	Bottle: 2 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished
169851	Omar Bach Rais	Bottle: 1 Label: Tag: Result: E. Coli: ≤ 4 /100ml. Total Coliform: ≤ 4 /100ml. Date: 2022-02-02 17:05:27 Location: 48.795621,2.451875	Finished

Measurement history search bar

Historical measurement

Generate report button

Show data graph button

Data Interface (Commands History)

Inline sensor data Measurement history **Commands history** System maintenance

Master Commands:

Log of commands:

220377	START MEASUREMENT 1,2,3,4,5,6	Omar Bach Rais	2022-02-01 15:51:48	2022-02-01 15:51:48	Response processed. Finished.	2022-02-01 15:53:12
220370	PING	Omar Bach Rais	2022-02-01 15:45:44	2022-02-01 15:45:44	Response processed. Finished.	2022-02-01 15:46:08
220369	INIT	System	2022-02-01 15:45:22	2022-02-01 15:45:22	Response processed. Finished.	2022-02-01 15:45:42
220355	PING	Omar Bach Rais	2022-02-01 12:32:55	2022-02-01 12:32:55	Response processed. Finished.	2022-02-01 12:33:24
220311	START MEASUREMENT 1,2,3,4,5,6	Patrick EA	2022-01-31 16:20:49	2022-01-31 16:20:49	Response processed. Finished.	2022-01-31 16:22:31
					Response processed.	

Start Date End Date

Data Interface (System Maintenance)

Inline sensor data Measurement history Commands history **System maintenance**

Device parameters

Device GPS:

Device ALERT : E. Coli Alert Total Coliform Alert

Measurement Type:

Default Calibration:

Tags: